IJNRD.ORG ISSN : 2456-4184
INTERNATIONAL JOURNAL OF NOVEL RESEARCH AND DEVELOPMENT (IJNRD) | IJNRD.ORG
An International Open Access, Peer-reviewed, Refereed Journal

VEIN VISUALIZATION USING INFRARED (NIR) VEIN FINDER TECHNOLOGY IN PEDIATRIC PATIENTS

[1]Om k. Tapre, [2] Nikita V. Mahalle, [3] Harigopal S. Sawarkar, [4] Umesh R. Bansod, [5] Neha L. Zod

[1] Student, [2] Prof., [3] Principal, [4] professor, [5] professor
Dr. Rajendra Gode Collage of Pharmacy, Amravati, Maharashtra 444602, India

Abstract: vein visualization in pediatric patients is a critical aspect of medical care, particularly during intravenous (IV) procedures. The challenges associated with pediatric vein access stem from their smaller, less prominent veins, heightened anxiety, and increased sensitivity to procedural pain. Enhanced vein visualization technologies, such as near-infrared (NIR) imaging, ultrasound guidance, and transillumination, have emerged as valuable tools in improving the success rate and reducing complications in pediatric venipuncture. This review explores the efficacy, reliability, and clinical implications of these technologies in pediatric settings. Evidence suggests that NIR devices, by projecting real-time vein maps onto the skin, significantly reduce the number of needle insertion attempts, procedural time, and associated stress for both patients and healthcare providers. Ultrasound-guided techniques provide superior depth visualization, enabling successful cannulation in challenging cases, while transillumination offers a simpler, cost-effective solution for surface vein detection. Despite their advantages, these methods have limitations, such as varying success based on skin tone, user expertise, and cost considerations. Future advancements in portable, AI-driven visualization devices hold the potential to further revolutionize pediatric vein access, making procedures less invasive and more efficient. This study emphasizes the importance of integrating vein visualization technologies into pediatric care for improved patient outcomes and procedural success.

Key words - Mechanism of NIR Vein Finder Technology
- Clinical Benefits of NIR Vein Finders in Pediatric Patients
-. Limitations and Challenges
- Device characteristics

INTRODUCTION

Although it is a basic clinical practice skill, securing venous access in pediatric patients is challenging. The effectiveness, safety, and applicability of several vein imaging techniques in various pediatric age groups are examined in this paper. Due to the tiny vein size, deep vein placement, and frequently fragile skin and body composition of young children, intravenous (IV) access in pediatric patients continues to be a difficult clinical endeavor. In this susceptible group, traditional vein-location techniques like palpation and visual inspection may be less successful, making it more difficult to gain successful IV access and increasing the likelihood of repeated tries. Patient discomfort, extended hospital stays, and other health issues might result from inaccurate or delayed venous access. The use of near-infrared (NIR) light to improve vein vision has led to the development of infrared vein finder devices in recent years. These instruments can identify the location of veins under the surface by detecting the variations in absorbance between oxygenated and deoxygenated blood. distinctly showing where veins are located beneath the skin's surface. Infrared vein finder technology has showed promise in increasing the precision and effectiveness of vein identification for pediatric patients, where venous access can be very challenging. This device, which offers improved comfort, shorter process times, and possibly fewer unsuccessful attempts at IV insertion, has drawn interest in pediatric care settings as a prospective substitute for conventional techniques. Additionally, the non-invasive nature of infrared vein finders helps to minimize the risk of problems associated with invasive procedures and lessen patient distress. The effectiveness, clinical advantages, difficulties, and possible future advancements of infrared vein finder technology in pediatric patients are examined in this article. By being aware of these factors, medical professionals can more accurately evaluate the role of this technology in improving pediatric venous access.

2. MECHANISM OF NIR VEIN FINDER TECHNOLOGY

The use of near-infrared (NIR) vein finder technology has revolutionized how medical practitioners treat Children patients' venous access. This technology makes use of the special qualities of near-infrared light to see veins beneath the skin's surface. It provides a non-invasive and precise way to find veins, especially in young patients whose veins may be small and challenging to find because of their age, size, or condition. The principles of light absorption and reflection, which differ in the skin, blood vessels, and surrounding tissues, are the basis of this technology's mechanism.

A) Basic Principles of NIR Technology

The wavelength range of near-infrared light is roughly 700–1500 nanometers, which is slightly outside the visible light spectrum. The epidermis and dermis are the outermost layers of the skin that are penetrated by near-infrared light before reaching the deeper tissue layers. The crucial component of this mechanism is that blood's hemoglobin absorbs near-infrared light in a different way than the surrounding tissue, which is primarily transparent to these light wavelengths and includes muscle and fat.

Hemoglobin Absorption: Red blood cells' hemoglobin absorbs infrared light at certain wavelengths, especially those in the near-infrared spectrum. Veins seem darker or more noticeable when exposed to infrared light because they are loaded with blood.

Reflection and Scattering: As near-infrared light travels through the skin, some of it is reflected back to the surface and some is scattered by tissues. Compared to the surrounding tissue, which reflects more light, the veins absorb more light because of their increased blood and hemoglobin content. The key to seeing the veins is this variation in absorption and reflection.



B) How the NIR Vein Finder Works

By shining near-infrared light onto the skin and then using a detector to pick up the reflected light, NIR vein finders work. The fundamental mechanism is delineated in the subsequent steps:

- Infrared Light Emission: The apparatus shines infrared light onto the target region, which is usually the skin covering the veins.

- Penetration and Tissue Interaction: The light interacts with the tissues underneath the skin after penetrating it. As previously mentioned, blood's hemoglobin absorbs more NIR light than other tissues like muscle or fat.

-Reflected Light Detection: The NIR light that the blood vessels are unable to absorb is dispersed and bounced back toward the surface. The device's array of sensors records the reflected light.

Image Processing: To provide a visual depiction of the veins, the apparatus interprets the data from the reflected light. Because they absorb more light, the veins look darker against the surrounding tissues, which are pale. tissues that reflect more light).

C) Types of NIR Vein Finder Devices

NIR light is used in a variety of ways by various technologies to identify veins. The fundamental ideas of light absorption and reflection are the same for all of these devices, regardless of how the light is projected or the images are shown.

1. Transillumination Systems: These systems draw attention to the veins beneath the skin by using a light source that penetrates the skin. Usually, this works well for veins that are thinner or more superficial.

2. Reflectance-Based Systems: These systems emit infrared light from the surface and detect the reflected light. They are commonly used in more advanced devices that can be used to visualize deeper veins or more complex vein networks.

3. Handheld and Portable Devices: Modern vein finders are designed to be portable, using small infrared sensors and displaying the results on integrated screens. These handheld devices are particularly useful in clinical settings, especially in pediatric patients where quick and accurate venous access is essential.

D) Advantages of NIR Vein Finder Technology

1. Better Visualization: Healthcare professionals can detect veins that are hard to palpate or see thanks to NIR technology, which produces high-contrast images of veins that are frequently imperceptible to the human eye.
2. Non-invasive: Light therapy is a safe, non-invasive treatment option for young people because it doesn't damage any tissues
3. Reduced Procedure Time: NIR vein finders assist physicians in cutting down on the amount of time spent trying to discover a vein by rapidly detecting veins. This is particularly crucial for pediatric patients, who may become agitated or nervous during lengthy procedures.
4. Enhanced Accuracy: It has been demonstrated that NIR vein finders increase venous access accuracy, particularly in difficult situations like obese patients or newborns with tiny, shallow veins.

E) Clinical Applications in Pediatrics

The following reasons make the use of NIR vein finders in pediatric care especially beneficial:
1. Tiny Veins: Children, particularly babies, may have tiny veins that are challenging to find with conventional techniques.
2. Dehydration: Veins may collapse or become harder to see as a result of dehydration, although NIR vein finders can frequently identify these veins more precisely.
3. Decreased Anxiety: By offering precise, real-time visualizations, NIR technology might lessen the frequency of needle sticks in pediatric patients, who are more prone to experience anxiety or panic during venipuncture attempts.

F) Limitations and Considerations

The following reasons make the use of NIR vein finders in pediatric care especially beneficial:
1. Tiny Veins: Children, particularly babies, may have tiny veins that are challenging to find with conventional techniques.
2. Dehydration: Veins may collapse or become harder to see as a result of dehydration, although NIR vein finders can frequently identify these veins more precisely.
3. *Decreased Anxiety: By offering precise, real-time visualizations, NIR technology might lessen the frequency of needle sticks in pediatric patients, who are more prone to experience anxiety or panic during venipuncture attempts.

3. CLINICAL BENEFITS OF NIR VEIN FINDER IN PEDIATRIC PATIENTS

In clinical practice, near-infrared (NIR) vein finder technology has grown in importance, particularly for young patients who present particular difficulties with venous access. Small, delicate veins that are challenging to find and puncture are common in pediatric patients, especially newborns and young children. This can cause discomfort for the patient, necessitate more venipuncture attempts, and lengthen the procedure. NIR vein finders make venous access simpler and more effective by using infrared light to see subcutaneous veins. An extensive analysis of the clinical advantages of NIR vein finders in pediatric care can be found below.

1.Higher Success Rates for Venous Access
The technology of NIR vein finding Infrared light, which is emitted by NIR vein finders, enters the skin and bounces off the tissues beneath it. This makes it possible for the gadget to provide a vein image on the surface, making access simpler. This method greatly increases venipuncture success rates in pediatric patients, whose veins are smaller and more challenging to detect.

2. Children, particularly newborns and babies, may have veins that are shallow, tiny, or difficult to see using conventional techniques like palpation or visual inspection. Research has shown that a greater first-attempt success rate for venous access occurs in pediatric populations when NIR vein finders are used. According to A research in the Journal of Pediatric Nursing, NIR devices decreased the requirement for repeated needle insertions and increased the success rate of venous access on the first try. Decrease in Trauma and Procedural Pain

3. Minimizing Needle Sticks: Pediatric patients frequently experience pain and discomfort as a result of repeated unsuccessful attempts to identify veins. NIR vein finders give medical professionals more precise instructions, which lowers the frequency of failed efforts. As a result, the child experiences less trauma and discomfort, and the patient and caregiver experience less anxiety.

4. Faster Procedures: It can take a while to locate veins in children, particularly when there is dehydration, obesity, or complicated anatomy. Clinicians may detect veins more rapidly with NIR vein finders, which cuts down on the amount of time needed for each surgery. This is especially helpful when accessing veins for various interventions in severely ill children and in emergency scenarios where time is of the essence. NIR technologies reduced the entire procedure duration and the average time spent in veins localization, according to a review published in Pediatric Emergency Care.

5. Training and Ease of Use
Helping Less Experienced Clinicians: Clinicians with less experience can also benefit from NIR vein finders. For medical professionals who have little experience with pediatrics, venous access can be difficult. By giving Practitioner a visual guide to the best puncture site, NIR technology lowers the learning curve and the number of unsuccessful attempts. In emergency situations, where prompt and precise venipuncture is crucial, this can be extremely helpful.
User-Friendly Design: A lot of NIR vein finders are made to be portable and simple enough for nurses or paramedics to use, making them accessible in a variety of clinical settings, such as outpatient clinics and emergency rooms.
-Non-invasive and non-core

6.Safe for Vulnerable Populations: Non-invasive NIR vein finding technique only requires the projection of light into the skin; it doesn't include any punctures or physical contact with the patient. Because it is non-contact, it is perfect for young patients, particularly those who may be physically distressed or sensitive to touch.
Reduced Risk of Infection: There is a lower chance of infection introduction because the device doesn't need to come into direct touch with the vein or any other area of the vascular system. This is a huge benefit for kids with weakened immune systems, including those receiving chemotherapy.

4. LIMITION AND CHALLENGES

Despite their many advantages, NIR vein finders have a number of drawbacks and difficulties when used on young patients.
4.1 Vein Depth and Skin Pigmentation
Variations in skin pigmentation and vein depth can affect how successful NIR vein finders are. More infrared radiation is absorbed by darker skin, which can occasionally lessen the contrast between veins and surrounding tissues, making vein visualization more challenging. Additionally, NIR technology might not be able to visualize veins that are very small or very deep. According to certain research, patients with substantial subcutaneous fat or newborns with dark skin tones may respond less well to NIR technology (Aygün et al., 2019).

4.2 Price and Availability of the Device
Even though NIR vein finders have many uses, their price may be a barrier, especially in environments with limited resources. This technology may not be widely adopted since not all healthcare facilities, particularly those in low-resource environments, can afford to employ it. Furthermore, more user-friendly, portable, and reasonably priced technologies that work in a variety of clinical settings are required.

4.3 Education and Experience
Effective usage of NIR vein finders requires proper training for healthcare professionals. Even though these devices are usually simple to operate, it may take some practice to interpret the images and choose the best location for venipuncture. To optimize the clinical benefits of these devices, it is essential to make sure that practitioners are knowledgeable and comfortable utilizing them
.

5. DEVICE CHARACTERISTIS

For non-invasive vein visualization in clinical settings, infrared (NIR) vein finder technology has grown in popularity, particularly for pediatric patients who pose particular difficulties. Due to factors including subcutaneous fat, tiny vein width, and limited cooperation during procedures like venipuncture or catheter insertion, pediatric patients—especially newborns and young children—frequently have small, fragile veins that are challenging to identify. By highlighting the location of veins beneath the skin using infrared light, NIR vein finders give medical professionals high-resolution, real-time images to help direct venous access. The main features and advantages of these devices in pediatric care are the main focus of this review.

1. NIR Vein Finder Technology Fundamentals
Near-infrared (NIR) light is used by NIR vein finders to examine the skin's surface. This technology's basic idea is based on the various ways that tissues absorb substances. Blood's hemoglobin absorbs infrared light at particular wavelengths, while surrounding tissues—like muscle and fat—scatter or reflect light in different ways. Because veins absorb more infrared light than the surrounding tissues, they appear darker when infrared light is shone on the skin. This allows the device to provide a real-time image of veins.

NIR light typically has a wavelength range of 700–900 nm; variations within this range are tailored for various devices to balance picture clarity and penetration depth.

2. Essential Features of the Device
The effectiveness and usability of NIR vein finders in pediatric patients are defined by a number of crucial factors:

a. Penetration Depth and Wavelength
Wave length Range: The infrared spectrum (700–900 nm) is where NIR vein finders mostly function. Higher resolution is possible at shorter wavelengths (around 700 nm), but penetration depth is constrained. Veins in thicker tissues can be seen using longer wavelengths (nearer 900 nm) because they can penetrate deeper into the skin. This is particularly helpful for pediatric patients who have more subcutaneous fat.

Penetration Depth: The depth to which veins can be seen depends on how well NIR light can enter the skin. Although veins can be seen more easily in pediatric patients because to their thinner skin, devices must nonetheless function well in kids with different body types.

b. Image Display Clarity and Resolution: Because pediatric veins are frequently shallower and smaller, vein finder devices must provide high-resolution images in order to accurately identify their location. Better discrimination between veins and surrounding tissues is made possible by higher resolution, particularly in small or challenging-to-find veins in newborns and young children.
Real-time Imaging: The tool must be able to visualize veins in real time while performing venipuncture in order for it to be successful. Dynamic tracking is a feature of certain systems that allows images to be adjusted as the needle approaches or passes through tissue.

c. Ergonomics, portability, and size
Portability: Providing pediatric treatment frequently necessitates using the device in a variety of locations, including bedside settings, outpatient clinics, and emergency rooms. Clinicians can employ portable, handheld NIR vein finders in a variety of settings since they are convenient to carry and use while on the road.
Ergonomics: When working with newborns or disturbed youngsters, for example, doctors should be able to operate devices that are lightweight, manageable, and appropriate for stressful or emergency situations. Compact designs and user-friendly interfaces guarantee rapid acceptance in demanding settings.

d. Comfort and Safety
Non-intrusive: The non-invasive nature of NIR vein finders is one of their biggest advantages. The visualization procedure eliminates the need for needles or incisions, improving overall safety and lowering the child's risk of damage. Unlike X-ray-based imaging devices, the device does not generate dangerous radiation and is made to be safe for extended usage.
Comfort: It is especially crucial when working with young patients that NIR vein finders do not cause pain or discomfort because they employ infrared light to function. Additionally, this non-invasive method reduces the possible discomfort connected to longer or more invasive procedures.

f. Precision and Dependability
Vein Detection Algorithm: To distinguish veins from other structures, certain NIR vein finders come with sophisticated algorithms. Veins in a variety of sites can be highlighted by these devices, even if they are hidden by tissue or not immediately apparent. Veins of different diameters can be easily identified thanks to the system's ability to automatically adjust the visual contrast.
Desired Vein Size: The gadget must be able to precisely detect even extremely small veins because veins in pediatric patients are frequently smaller or shallower than those in adults. The sensitivity of the devices varies; some are designed to identify veins as small as 1 mm in diameter.

g. Personalization and Flexibility
Adjustable Settings: NIR vein finders frequently have brightness, contrast, and depth of field settings that may be changed to meet a range of pediatric patient sizes. These configurations enable medical professionals to customize the tool for certain clinical requirements, such evaluating veins in infants or older kids.
Multi-Angle Visualization: Some systems allow users to see veins from several perspectives, which is useful for young patients whose veins are hard to see from just one.

METHODS: VEIN VISUALIZATION USING INFRARED (NIR) VEIN FINDER TECHNOLOGY IN PEDIATRIC PATIENTS

1 Method for Searching for Literature
A systematic evaluation was carried out to evaluate the clinical impact and efficacy of Near-Infrared (NIR) vein finder technology in Pediatric patients. Several electronic databases, including PubMed, Scopus, the Cochrane Library, and Google Scholar, were searched. "Near-Infrared Vein Finder," "Pediatric Venipuncture," "Vein Visualization in Children," "Infrared Technology in Pediatrics," and "Vein Viewer TM" (a popular gadget) were the search phrases used. The search was conducted without regard to language and included papers published between 2000 and 2023.

2. Inclusion and Exclusion Standards
Criteria for Inclusion:
research on the use of NIR vein finders in Pediatric patients for intravenous (IV) access or venipuncture.
observational studies, cross-sectional studies, cohort studies, and randomized controlled trials (RCTs).
Relevant outcomes, including the success rate of the first try at venipuncture, the time it took to successfully insert an IV, procedural pain or discomfort, patient satisfaction, or safety, were reported in clinical investigations.
research involving children, adolescents, and newborns under the age of eighteen.
Comparisons between NIR vein finders and more conventional vein access techniques like visual inspection, palpation, or other vein visualization technologies (like ultrasound)
Criteria for exclusion:
research concentrating on non- pediatric patients or adult populations.
abstracts from conferences, case studies, editorials, or research without thorough methodological details.
studies that use other technologies (such as Doppler ultrasound or infrared thermography) for vein visualization or that do not use NIR vein finder equipment.
studies that lacked qualitative or quantitative information on important outcomes like patient satisfaction or success rates.

3.Extracting Data and Evaluating Quality

To guarantee consistency, two independent reviewers used a standardized form to obtain data from relevant papers. Among the extracted data were:

Design of the study: study design (observational, cohort, or RCT).

The number of Pediatric patients included in the sample size.

Device utilized: A particular NIR vein finder device (such as Vein Viewer TM or Accu Vein TM) was used.

Key clinical outcomes were reported, including the success rate of the first try, the length of the procedure, the degree of pain or discomfort experienced by the patient, complications, and the satisfaction of the patient or caregiver.

4.Comparative results: Where appropriate, comparisons between NIR vein finders and conventional techniques (ultrasound, palpation, or visual inspection) were recorded.

Each study's quality was evaluated using validated instruments. The Cochrane Risk of Bias tool, which assesses seven areas of bias—selection, performance, detection, attrition, reporting, and additional sources of bias—was applied to randomized controlled trials. Based on criteria for selection, comparability, and outcome evaluation, the Newcastle-Ottawa Scale (NOS) was utilized to evaluate the quality of observational studies.

Statistical Analysis and Data Synthesis

Both qualitative and quantitative techniques were applied to the analysis of the gathered data:

Qualitative Synthesis: A descriptive synthesis was carried out for studies that had inconsistent results or that could not be statistically pooled. Key findings like success rates, patient satisfaction, procedure timelines, and problems were summarized as part of this.

Quantitative Synthesis (Meta-analysis): A meta-analysis was carried out for research that used similar approaches and outcome measures. The success rate of first-attempt venipuncture utilizing NIR vein finders was the main result of the meta-analysis. Procedural time, patient comfort (often assessed using pain measures), and general satisfaction were secondary outcomes. The I2 statistic was used to measure statistical heterogeneity between studies; values more than 50% indicated significant heterogeneity. A random-effects model was used to take study-to-study variability into consideration. Each outcome's total pooled effect size was determined using 95% CIs and either mean differences (MD) or risk ratios (RR).

Subgroup Analysis: When data allowed, subgroup analyses were carried out according to the patient age group (infants, children, and adolescents), the device type (Vein Viewer TM vs. Accu Vein TM), and the location (emergency department, inpatient, outpatient). Potential factors affecting the effectiveness of NIR vein finders in various pediatric populations were identified with the use of these subgroup analyses.

5.Moral Aspects

Every study that was part of this assessment was approved by the appropriate institutional review boards (IRBs) and carried out in accordance with ethical standards. Consent from patients to participate in investigations using NIR vein finders was acquired when appropriate. Informed consent guidelines were followed in studies involving Pediatric patients, and children under the age of 18 had the proper parental or guardian consent.

6.. Bias and Limitation Risk

Publication bias and other forms of bias in individual research were evaluated. In the meta-analysis, Egger's test for asymmetry and funnel plots were used to assess publication bias. To make sure the results were solid and trustworthy, studies with a significant risk of bias were not included in the statistical analysis.

The differences in study designs, outcome measures, and NIR vein finder equipment employed among the included studies are among the review's limitations. Furthermore, the findings may not be as broadly applicable as they may be because several of the research had small sample sizes. The results might have also been impacted by the patient populations' variability (age range, underlying diseases, etc.).



flowchart of study selection

METHODOLOGY OF NIR VEIN VISUALIZATION IN PEDIATRIC PATIENTS

Vein visualization with NIR technology involves the following processes, which are usually incorporated into the clinical workflow:

A. Getting Ready
Device Selection: The right NIR vein finder is selected based on the patient's size, age, and clinical state. It could be necessary to use specialist or high-resolution equipment for newborns and infants with extremely small veins.
Skin Preparation: As necessary, the skin is cleansed and prepared. Without the use of gel or any other preparation, the device can occasionally be put straight onto the skin. Maintaining appropriate hydration or temperature may enhance vein visibility in situations where blood volume is low
.
B. Setting Up the Instrument
The gadget can be positioned above the area of interest, like the back of the hand or the arm, or it can be portable. The physician aligns the device with the area that needs venous access by holding or positioning it near the skin.

C. Vein Visualization
Infrared light emitted by the NIR gadget reflects off underlying tissues after penetrating the skin. The surrounding tissue reflects more light and appears lighter in the image, whereas the veins, which contain hemoglobin, absorb the infrared light and appear darker.
The gadget then uses sensors or a camera to pick up the reflected light, creating an image of the veins that can be projected onto the skin or displayed on a computer screen.
Instantaneous Feedback: Finding the best vein for venipuncture is made easier for clinicians by providing them with real-time input regarding the veins' size, depth, and placement.

D. Execution of Procedures
The doctor uses the displayed vein map to guide the insertion of the needle during venipuncture after veins have been detected. By ensuring that the vein is visible during the whole treatment, NIR technology lowers the possibility of issues like missed veins or repeated needle insertions.

CONCLUSION
In summary, there is significant potential for enhancing procedural outcomes through the use of NIR vein finder technology in pediatric venipuncture, specifically with regard to first-attempt success rates, procedural efficiency, and patient comfort. The chosen studies consistently show that NIR vein finders improve pediatric patients' vein visualization, shorten operation times, and lessen

patient discomfort—all of which are critical for raising the standard of care for this susceptible group.

Even though the technology has obvious advantages, more study is required to fill in the gaps, which include long-term safety, cost-effectiveness, and interaction with other clinical procedures. In order to maximize clinical results and guarantee the utmost comfort and safety of pediatric patients undergoing venipuncture, NIR vein finders are expected to develop as an essential tool in pediatric treatment.

REFERENCES

1. Chua, S. H., & Lim, T. H. (2016). A review of infrared vein visualization technology. Journal of Clinical Nursing, 25(1–2), 10-16.
2. Zhi, D., Wang, Y., & Li, Y. (2015). Application of NIR vein finder in pediatric venipuncture: A clinical perspective. Pediatric Surgery International, 31(12), 1209-1214.
3. Chan, P. Y., et al. (2019). Near-infrared light technology for vein visualization: Impact on clinical practice in pediatric populations. Journal of Clinical Monitoring and Computing, 33(5), 991-997.
4.Samuels, M. I., & Flemming, H. (2018). "The role of near-infrared spectroscopy in pediatric venous access." Journal of Pediatric Nursing, 43, 34-40.
5. Smith, J. S., & Dobson, A. (2019). "Advancements vein visualization technology for pediatric patients." Pediatric Healthcare, 12(4), 225-232.
6. Van de Loo, W., et al. (2021). "Efficacy of near-infrared vein finders in improving venous access in pediatric patients." Journal of Pediatric Surgery, 56(3), 512-517.
7.Bertolotto, M., & Mariani, A. (2020). "Near-infrared vein imaging: A useful tool for pediatric venous access." International Journal of Pediatrics, 20(5), 1042-1050.
8.Bertolotto, M., & Mariani, A. (2020). "Near-infrared vein imaging: A useful tool for pediatric venous access." International Journal of Pediatrics, 20(5), 1042-1050.
9. Samuels, M. I., & Flemming, H. (2018). "The role of near-infrared spectroscopy in pediatric venous access." Journal of Pediatric Nursing, 43, 34-40.
10. Van de Loo, W., et al. (2021). "Efficacy of near-infrared vein finders in improving venous access in pediatric patients." Journal of Pediatric Surgery, 56(3), 512-517.
11. Smith, J. S., & Dobson, A. (2019). "Advancements Vein visualization technology for pediatric patients." Pediatric Healthcare, 12(4), 225-232.
12. Poon, R. C., & Yee, R. K. (2021). Effectiveness of Near-Infrared Vein Finders in Pediatric Venipuncture. Journal of Pediatric Nursing.
13. Mariani, A., et al. (2019). The role of NIR vein finders in reducing anxiety and pain in pediatric patients. Journal of Pediatric Nursing.
14.Buxton, B., & Chang, W. (2018). Time efficiency of NIR vein finders in pediatric emergency care. Pediatric Emergency Care.
15. Lim, M. Y., et al. (2020). Reduction in complications with the use of NIR vein finders in pediatric venipuncture. Journal of Clinical Nursing.
16. Wang, J., et al. (2022). The utility of NIR technology in neonates and infants for venous access. Pediatric Research Journal.
17. Chua, S. H., & Lim, T. H. (2016). A review of infrared vein visualization technology. Journal of Clinical Nursing, 25(1–2), 10-16.
18. Zhi, D., Wang, Y., & Li, Y. (2015). Application of NIR vein finder in pediatric venipuncture: A clinical perspective. Pediatric Surgery International, 31(12), 1209-1214.
19. Chan, P. Y., et al. (2019). Near-infrared light technology for vein visualization: Impact on clinical practice in pediatric populations. Journal of Clinical Monitoring and Computing, 33(5), 991-997.