IJNRD.ORG ISSN : 2456-4184
INTERNATIONAL JOURNAL OF NOVEL RESEARCH AND DEVELOPMENT (IJNRD) | IJNRD.ORG
An International Open Access, Peer-reviewed, Refereed Journal

High Performance Liquid Chromatography (HPLC) : A Short Review

Pansare Akanksha , Nibe Vaishnavi Pravara Rural College of Pharmacy , Loni

• Abstract :

One essential analytical method that is frequently used in both qualitative and quantitative analysis is high performance liquid chromatography (HPLC), which is especially useful for estimating biological and pharmaceutical substances. One of the most sophisticated chromatographic techniques, high-performance liquid chromatography (HPLC) is acknowledged for its adaptability, safety, dependability, and speed. It is essential to the assurance and quality control of medicinal constituents. The goal of this article is to provide a thorough overview of all the aspects of HPLC, including its foundational ideas, varieties, equipment, and wide range of applications in numerous industries. Comprehending these facets allows one to recognize the importance of HPLC in contemporary analytical science and its roles in guaranteeing the security and effectiveness of medications.

These days, a wide range of industries, including pharmaceuticals, biotechnology, environmental analysis, polymer research, and the food industry, heavily rely on HPLC for separations and purifications. Its accuracy and effectiveness in dissecting intricate combinations into their component parts account for its wide range of applications. A tiny volume of liquid sample is injected into the mobile phase—a continuously running stream of liquid—during the procedure. The sample is transported by this mobile phase via a column that is full of the stationary phase's densely packed particles. The components of the sample separate according to different physical or chemical properties as a result of their varied interactions with the stationary phase as it passes through the column. Because HPLC can precisely separate and purify molecules, it is an invaluable instrument for quality control and analysis of a wide range of substances in both industry and research.

• Keywords :

Introduction , Principle , Types , Overview, Basic HPLC Components , Apparatus , Separation , How to read chromatogram and Applications

• Introduction :

The domains of analytical chemistry and biochemistry frequently employ high-performance liquid chromatography (HPLC), commonly referred to as high-pressure liquid chromatography. HPLC is a particular type of column chromatography. This method is used to extract, recognize, and measure the active ingredients in a combination. Through the use of high pressure and a column packed with a solid adsorbent material, high performance liquid chromatography (HPLC) enables accurate and effective separation of the constituents according to their interactions with the stationary phase. This technique is widely applied in many fields, including as clinical diagnostics, food safety, environmental monitoring, pharmaceutical research, and mixture analysis. It is especially useful for accurately identifying and quantifying specific chemicals in complicated mixtures.

High-performance liquid chromatography (HPLC) retention times can differ greatly based on the interactions that occur between the stationary phase, the molecules under investigation, and the solvent or solvents that are

being utilized. Since the mobile phase transports the sample, which was added to the system in a small volume, through the column, these interactions are crucial to the separation process. Depending on their unique
characteristics and affinities for the stationary phase, certain components of the sample are retarded, or slowed down, as it moves forward due to certain chemical or physical interactions with it.

High-performance liquid chromatography (HPLC) retention times can differ greatly based on the interactions that occur between the stationary phase, the molecules under investigation, and the solvent or solvents that are being utilized. Since the mobile phase transports the sample, which was added to the system in a small volume, through the column, these interactions are crucial to the separation process. Depending on their unique
characteristics and affinities for the stationary phase, certain components of the sample are retarded, or slowed down, as it moves forward due to certain chemical or physical interactions with it.

Every analyte experiences a different degree of retardation depending on the kind of analyte and the makeup of the stationary and mobile phases. The retention time of a specific analyte is the moment at which it leaves the column and is discovered. Retention time is a crucial HPLC characteristic that aids in the identification and
measurement of the various components in the sample.

The properties of the stationary phase and the analytes under investigation dictate the selection of solvents, additives, and the specific gradient utilized in HPLC analysis. Achieving accurate results and optimizing the separation depends on a number of factors, including the composition of the stationary phase, the polarity, molecular weight, and chemical characteristics of the analysts

• PRINCIPLE OF HPLC :

In a separation column with a stationary and a mobile phase, the purification occurs. In a separation column, the stationary phase is a granular substance containing tiny porous particles. On the other hand, a solvent or solvent mixture known as the mobile phase is pushed through the separation process under intense pressure.

A small volume of liquid sample is injected into a tube filled with tiny particles (3 to 5 microns in diameter, called the stationary phase) in HPLC, a separation technique. The liquid (mobile phase) is forced through the column by high pressure provided by a pump, moving individual components of the sample down the packed tube (column).



Principle of HPLC

- **Types of HPLC :**

1) Normal Phase HPLC :

In normal phase HPLC, a polar stationary phase, like silica, is added to the column. In contrast, the mobile phase is often composed of organic solvents like hexane and is non-polar. This type of HPLC can be used to separate chiral molecules, geometric isomers, cis-trans isomers, and water-sensitive substances. The separation is based on the compounds' differing affinities for the polar stationary phase and the non-polar mobile phase. Polar chemicals interact with the stationary phase more strongly and elute more slowly.

2) Reverse Phase HPLC:

Reverse phase HPLC is the most widely used type. In this instance, the organic solvent that is miscible, such as methanol or acetonitrile, is combined with water to form the polar mobile phase. The non-polar stationary
phase is often composed of a hydrophobic substance such as C18 (octadecylsilane). This design makes Reverse Phase HPLC flexible enough to handle a wide range of compounds, including polar, non-polar, ionizable, and ionic samples. Because of their weaker connections with the non-polar stationary phase, less polar molecules elute more quickly, which is what propels the separation process brought on by hydrophobic interactions.

3) Ion Exchange HPLC:

Analytes with the opposite charge interact with ionic groups in the stationary phase of an ion exchange HPLC. The mobile phase is typically a buffer solution that controls the pH and ionic strength of the surrounding medium to aid in the separation of ionic species. This technique is particularly effective for isolating and purifying anions and cations, such as proteins, nucleotides, peptides, and amino acids, based on their charge.
The intensity of the interaction between the analyte and the stationary phase controls the order of elution; weaker interactions result in earlier elution .

4) Size Exclusion HPLC :

Size Exclusion Another name for HPLC is gel permeability chromatography (GPC), a method that separates molecules based on size. The stationary phase molecules are arranged in a porous media based on their ability to enter the pores of the packing material. Large molecules flow through the column more quickly and elute first because they are excluded from the pores. Because they can fit through the pores, smaller molecules go farther through the column and elute later. When looking at proteins, polymers, and other macromolecules to find out more about their size distribution, this method is particularly useful.

5) Bio affinity HPLC :

Based on a specific, reversible protein-ligand interaction for separation. Proteins that interact with columnbound ligands are retained when ligands are covalently attached to solid supports in a bio-affinity matrix. There are two methods for eluting proteins that are attached to a

bioaffinity column:

- Biospecific elution: In order to compete with ligand that is attached to the column, free ligand is introduced to the elution solution.
- Aspecific elution: a shift in pH, salt, or other factors that decreases the interaction between the interaction protein and the column-bound substrate.

• Overview of HPLC :

The advanced analytical technique known as High Performance Liquid Chromatography, or HPLC, is used to identify, quantify, and separate the components of a mixture. The word "chromatography" refers to a broader class of techniques for breaking down mixtures into their component elements based on differences in how the mixtures interact with a stationary phase and a mobile phase.

In chromatography, the outcome or graphic representation of the separation process is called a "chromatogram." This picture usually shows the components of the mixture separated over time, with various compounds represented by different peaks.

A "chromatograph" is a piece of equipment used in chromatography. The chromatograph, a cutting-edge tool in high-performance liquid chromatography (HPLC), efficiently separates substances by applying high pressure to a liquid mobile phase that is packed into a stationary phase-packed column.

The solvent used to separate a liquid sample's constituent parts for HPLC analysis is known as the mobile phase. The mobile phase is given to the detector first, then to a separation column (also known as the stationary phase), at a constant flow rate controlled by the solvent delivery pump.



• Basic HPLC System Components :

◆ **Solvent Degasser** : An essential part of the HPLC system, the solvent degasser is made to extract dissolved gases from the solvent, like nitrogen and oxygen. Bubbles created by dissolved gases in the pump and detector might cause baseline noise and anomalies in the chromatographic results. The degasser improves the accuracy and dependability of the analysis by efficiently getting rid of these gases and ensuring a steady and consistent flow of the mobile phase.

◆ **HPLC Pump :** Delivering the mobile phase through the system at a precise and constant flow rate is the job of the HPLC pump. It must be able to produce high pressures—often as high as several thousand psi—in order to push the solvent through the column's densely packed particles. The pump can also precisely combine various solvents in a proportionate manner, enabling gradient elution—an essential process for separating complex combinations.

◆ **Autosampler :** The HPLC system's sample injection procedure is automated by the autosampler. It takes exact sample volumes out of vials and injects them into the stream of the mobile phase. This automation lowers the possibility of human error while increasing throughput. Repeatable results require high-throughput analysis and uniform sample injection volumes, which autosamplers can handle several samples in sequence to enable**.**

◆ **Detector :** An essential part of detecting and quantifying the analytes as they elute from the HPLC column is the detector. Mass spectrometers, UV-Vis absorbance detectors, and fluorescence detectors are examples of common detector types. Analytes' chemical or

physical characteristics, like ionization or absorbance, are transformed into an electronic signal by the detector. The software of the device then processes this signal to create a chromatogram, in which the peaks represent various analytes.

◆ **Column Oven :** For repeatable and effective separations, the HPLC column must be kept at a constant temperature, which is maintained by the column oven. The viscosity of the mobile phase, the interaction of the analytes with the stationary phase, and the total retention duration of the analytes can all be impacted by temperature control. The column oven guarantees reliable separations and optimal chromatographic performance by maintaining the column at a constant temperature.

• Apparatus of HPLC :

We have now covered the "Basic Overview of the HPLC process" and its workings. More specifically, a solvent delivery pump, a degassing unit, a sample injector, a column oven, a detector, and a data processor are some of the parts that make up an HPLC.



Regarding HPLC, the mobile phase is delivered by the pump at a regulated flow rate (a). Under our normal atmospheric pressure, air can dissolve readily in the mobile phase. Problems include flow rate changes and baseline noise/drift may arise if the mobile phase, which contains air bubbles, enters the delivery pump. By eliminating air bubbles in the mobile phase, the degassing device (b) helps avoid this problem. Following the extraction of the dissolved air, the mobile phase is introduced into the column. Next, a standard solution or sample solution is added via the sample injector to the mobile phase (c). Variations in temperature may have an impact on how well the compounds separate in the column. To maintain a constant temperature, the column is placed in a column oven (d). Detectors downstream of the column (e) pick up compounds that elute from the column. To identify and measure the chemicals, a workstation processes the detector's signal to create a chromatogram (f).

- **HPLC Separation :**



The highly developed and adaptable analytical method known as High-Performance Liquid Chromatography (HPLC) is widely applied in a number of industries, including biochemical research, environmental testing, food and beverage analysis, and pharmaceuticals. HPLC is an essential instrument for quality control, purity evaluation, and the study of complex biochemical processes since it is made to separate, identify, and measure the different components inside complex mixtures.

Differential migration of substances through a column filled with a stationary phase while being transported by a mobile phase is the foundation of high-performance liquid chromatography (HPLC). The distinct chemical and physical interactions that every component has with the stationary phase and the mobile phase have an impact on this differential migration. Because of this, the chemicals in a mixture separate as they go through the column at various speeds, making it possible to identify and quantify each one of them separately.

Highly sensitive detectors, including UV-Vis, fluorescence, or mass spectrometry, are built into HPLC systems to provide precise and accurate measurements of the eluting chemicals. The information produced can be utilized to identify unidentified components, calculate known compound concentrations, and even evaluate sample purity. Furthermore, because of HPLC's adaptability, many parameters can be optimized to get the best possible separation and resolution of the target chemicals. These parameters include the temperature, the kind of stationary phase, the composition of the mobile phase, and the flow rate.

When samples are prepared properly, HPLC can handle a variety of sample types, including liquids, solids, and gases. This is one of its main advantages. Because of its great sensitivity and resolution as well as its adaptability, HPLC is a vital instrument in disciplines requiring exacting analytical precision. For instance, HPLC is essential in pharmaceutical research for the investigation of medication formulations, the identification of active components, and the stability monitoring of drugs. It is used in environmental science to identify and measure toxins and pollutants in samples of water, soil, and air.

Furthermore, improvements in HPLC technology, such as the creation of ultra-high-performance liquid chromatography (UHPLC), have improved analysis speed, resolution, and sensitivity even further, making it possible to complete even more complex separations in less time. These developments keep HPLC relevant in the changing field of analytical chemistry by extending its capabilities and uses.

- **How to Interpret a Chromatogram:**

A plot produced by chromatography is referred to as a "chromatogram". The chromatogram is a twodimensional figure where the horizontal axis represents the analysis time and the vertical axis shows

concentration as a function of detector signal strength. The resulting plot is not so much triangular in shape as it is bell-shaped. We refer to this form as a "peak."

The time gap between the sample injection point and the peak's apex is known as the retention time (tR). The dead time (t0) is the amount of time needed for non-retained compounds—that is, chemicals without any interaction for the stationary phase—to go from the injector to the detector.



Peak area (A), shown in light blue, is the region bounded by the peak and baseline. Peak height (h) is the vertical distance between a peak's apex and the baseline. The components of a sample will be analyzed both qualitatively and quantitatively using these results.

• Application of HPLC :

The HPLC has developed into a universally applicable method so that it finds its use in almost all areas of chemistry, biochemistry, and pharmacy.

- Analysis of drugs
- Analysis of synthetic polymers
- Analysis of pollutants in environmental analytics
- Determination of drugs in biological matrices
- Isolation of valuable products

- Product purity and quality control of in0dustrial products and fine chemicals

• Advantages of HPLC :

1. Speed
2. Efficiency.
3. Accuracy.
4. Versatile and extremely precise when it comes to identifying and quantifying chemical components

• Limitations of HPLC :

1. Cost: Despite its advantages, HPLC can be costly, requiring large quantities of expensive organics.
2. Complexity
3. HPLC does have low sensitivity for certain compounds, and some cannot be detected as they are irreversibly adsorbed.

• conclusion :

The review as a whole leads to the conclusion that HPLC is a flexible and reliable chromatographic method, particularly useful for the estimation of pharmaceutical goods. Because of its accuracy and dependability, it is a vital instrument in pharmaceutical analysis, where precise measurement of active substances is essential for regulatory compliance, quality control, and medication development.

Because HPLC can estimate active compounds both quantitatively and qualitatively, it is useful not only in the pharmaceutical industry but also in many other sectors. It is used in environmental science to identify and
measure contaminants in order to guarantee environmental security and adherence to legal requirements. HPLC is used in the food and beverage industry to verify product quality and safety, identify impurities, and analyze nutritional content. HPLC also helps with the study of metabolites, biomarkers, and other physiologically significant substances in clinical and biochemical research, which advances our knowledge of and ability to
treat a wide range of disorders .

The extensive applicability of HPLC is further demonstrated by its versatility in handling a large range of solvents and stationary phases, as well as its capacity to adapt to a variety of sample types, from simple molecules to complex mixtures. From routine quality control to advanced research and development, the approach is recommended due to its high sensitivity, precision, and capacity to differentiate closely related molecules .

• Reference :

1) Effects of high pressures in liquid chromatography, Martin M., Guiochon, G. 2005; J. Chromatogr. A; (12)7: 16–38

2) Siff rin C, Spoendlin C, Gerber F, Krummen M, Potgeter H, Roth A. A practical look at fast reversed- phase high-performance liquid chromatography in pharmaceutical development and manufacture under current good manufacturing practices, employing 3 microm particle packed columns and monolithic columns. Pubmed: 1546913. J Chromatogr A. 2004 May 21;1036(2):127-33. doi: 10.1016/j.chroma.2004.02.056.

3) Xiang Y, Liu Y, and Lee ML. Utilizing higher temperature for ultrahigh pressure liquid chromatography. 2006 Feb 3;11104(1-2):198-202; J Chromatogr A. 10.1016/j.chroma.2005.11.118 is the doi. PMID: 16376355.

4) Bergh JJ, Breytenbach JC. High-performance liquid chromatographic analysis of trimethoprim in pharmaceuticals: stability-indicating results.

PMID: 3558640; J Chromatogr. 1987 Jan 30;387:528–31. doi: 10.1016/s0021-9673(01)94565-0.

5. United States Pharmacopoeia, 27th ed., Rockville, MD: The USP Convention Inc., 2014.

6) Albero MI, Sanchez-Pedreno C, Garcia MS, and Rodenas V. Pharmaceutical flow-injection spectrophotometric measurement of sulphathiazole or frusemide. 1997; 15: 1687-1693 Journal of Pharmacy and Biomedical Analyst.

7. Abidi SL. Phosphatidic acids and related polar lipids: high-performance liquid chromatography. Chromatography Journal. 1991; 587: 193-203.

8) The European Pharmacopoeia, Council of Europe, Strasbourg, 2002, fourth ed.

9) Rice, J. J., Bleiberg, B., Muhammad, N., and Hanna, S. A. Lauback, R. G. 1984. Reversed-phase Ion-pair Chromatography for the Particular High-performance Liquidchromatographic Determination of Ampicillin in Bulks, Injectables, Capsules, and Oral Suspensions. 1984; 7(6): 1243–1265 in J. Liq. Chromatogr.

10) Wiklund A. E., Dag B., Brita S. Assessing toxicity by the use of sediment extracts and entire sediments. Bulletin on Marine Pollution (2005); 50(6): 660–667.

11) Toxicity identification assessment (TIE) of pore water of polluted marine sediments collected from Hong Kong waters by Kwok Y. C., Hsieh D. P. H., and Wong P. K. Bulletin on Marine Pollution, 51(8–12), 2005, pp. 1085–1091.

12) The use of toxicity identification evaluation methodologies on wastewaters and sludge from a municipal sewage treatment works with industrial inputs was carried out by Hongxia Yu, Jing C, Cui Y, Shang H, Ding Z, and Jin H in 2013. Environmental Safety and Ecotoxicology, 57(3), 426-430, 2004.

13) Ayerton J. High-pressure liquid chromatographic assay of ceftazidime in biological fluids. Chemother J. Antimicrob. 1981; 8: 227–231.

14) High-pressure liquid chromatographic test of sulbactam in plasma, urine, and tissue; Bowden R.E., Madsen P.O. Agents Chemother Antimicrob. 1986; 30: 31-233.

15) Haginaka J., Yasuda H., Uno T., Nkagawa T. High-performance liquid chromatography for alkaline degradation and detection. Bull. Chem. Pharm. 1984; 32: 2752-2758